



## **National Accreditation Board for Testing and Calibration Laboratories**

(A Constituent Board of Quality Council of India)



### **CERTIFICATE OF ACCREDITATION**

# MANAS LIQUID FLOW CALIBRATION LABORATORY (A DIVISION OF MANAS MICROSYSTEMS PVT. LTD.)

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

EL-54, Electronics Zone, MIDC, Bhosari, Pune, Maharashtra

in the field of

**CALIBRATION** 

**Certificate Number** 

CC-2832

**Issue Date** 

30/08/2018

Valid Until

29/08/2020

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Signed for and on behalf of NABL



Anil Relia

Chief Executive Officer





### National Accreditation Board for Testing and Calibration Laboratories

Q QCI

(A Constituent Board of Quality Council of India)

#### SCOPE OF ACCREDITATION

Laboratory

Manas Liquid Flow Calibration Laboratory (A Division of Manas Microsystems Pvt. Ltd.), EL-54, Electronics Zone, MIDC, Bhosari,

Pune, Maharashtra

**Accreditation Standard** 

ISO/IEC 17025: 2005

**Certificate Number** 

CC-2832

Page

1 of 1

Validity

30.08.2018 to 29.08.2020

Last Amended on -

SI.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability (±)	Remarks
# 1		FLUID FLOW	/ CALIBRATION	
I.	FLOW MEASURING D	EVICES		
1.	Volumetric Flow Rate (Water) <sup>\$</sup>	0.2 m <sup>3</sup> /h to 40 m <sup>3</sup> /h 40 m <sup>3</sup> /h to 150 m <sup>3</sup> /h 150 m <sup>3</sup> /h to 250 m <sup>3</sup> /h	0.3 % 0.3 % 0.3 %	Using Weighing Systems (500 kg, 2000 kg and 5000 kg), Flow Averaging Unit, Timer, Density Measuring System and Diverter by Gravimetric Method as per ISO 4185

Measurement Capability is expressed as an uncertainty (±) at a confidence probability of 95%

<sup>\$</sup>Only in Permanent Laboratory

Dheeraj Chawla Convenor

They have

A Ros

Avijit Das Program Manager